

A-dec® Simulator Maintenance

After Each Procedure (Student)

- ☐ Clean and lubricate handpieces (sterilize as appropriate).
- ☐ Clean the simulator by wiping all surfaces with mild dishwashing liquid and water.
- ☐ Flush vacuum lines, including the mask drain.
- ☐ Empty and clean the vacuum collection jar (AVS only).
- ☐ Add A-dec ICX® to each water bottle refill (if optional water bottle used)*.

Weekly, or As Needed (Student or Maintenance Staff)

Date/Time

- ☐ Clean HVE and saliva ejector valves.
- ☐ Inspect and maintain solids collector (central vacuum option only), twice per week.
- ☐ Maintain oil collector.
- ☐ Lubricate the handpiece chucks.
- ☐ Clean the LED dental light shield with mild dishwashing liquid and water. Use 2" x 2" gauze or soft cloth to avoid scratching the shield.
- ☐ Replace dental light Medicom® SafeShield™ barrier, if applicable.
- ☐ Clean chair and stool upholstery with mild dishwashing liquid and water.

Monthly (Student or Maintenance Staff)

Date/Time

- ☐ Inspect all handpiece and motor O-rings; replace if worn/damaged.
- ☐ Lubricate HVE and saliva ejector O-rings.
- ☐ Test dental unit water CFU level and shock waterlines with A-dec ICX Renew® or ICX Restore™, as needed*.

Yearly (Maintenance Staff)

Date/Time

- ☐ Replace and lubricate the syringe O-rings and button-valve assemblies.
- ☐ Replace and lubricate the HVE and saliva ejector O-rings.
- ☐ Replace all handpiece and motor O-rings.
- ☐ Replace air and water filters in the utility area.
- ☐ Lubricate water bottle O-rings.

*For detailed waterline maintenance instructions, refer to the Waterline Maintenance Guide (p/n 85.0983.00) that came with your equipment. For quick online access to this document, scan this code or go to www.a-dec.com/resource center.





 A-dec Headquarters

2601 Crestview Drive
Newberg, Oregon 97132
United States
Tel: 1.800.547.1883 within USA/CAN
Tel: +1.503.538.7478 outside USA/CAN
www.a-dec.com



86.0703.00 Rev B
Date of Issue 2023-10-17
Copyright 2023 A-dec, Inc.
All rights reserved.

IFUself9